

# Certificate of Analysis

## Product Description:

Part Number: **ICP-MS-68B**  
**Solution A**  
 Lot Number: **SAMPLE**  
 Matrix: 4% HNO<sub>3</sub>  
 Purity: 99.98% - 99.9999%

## Certified Values:

Element	(mg/L)	SRM ID	SRM Lot#	Element	(mg/L)	SRM ID	SRM Lot#
Al	100.0 ± 0.6	*		Lu	100.0 ± 0.6	3130a	100503
As	100 ± 1	3103a	100818	Mg	100.0 ± 0.6	3131a	140110
Ba	100.0 ± 0.6	*		Mn	100.0 ± 1.5	3132	050429
Be	100 ± 1	3105a	090514	Nd	100 ± 1	3135a	992803
Bi	100.0 ± 0.6	3106	991212	Ni	100.0 ± 0.6	3136	120619
B	100 ± 1	*		P	100.0 ± 1.5	3139a	060717
Cd	100.0 ± 0.6	3108	130116	K	100 ± 1	3141a	140813
Ca	100.0 ± 0.6	3109a	130213	Pr	100 ± 1	3142a	990501
Ce	100.0 ± 0.6	3110	090504	Re	100 ± 1	3143	010816
Cs	100 ± 2	3111a	130228	Rb	100 ± 1	3145a	891203
Cr	100.0 ± 0.6	3112a	030730	Sm	100.0 ± 0.6	3147a	892911
Co	100.0 ± 0.6	3113	000630	Sc	100 ± 1	3148a	100701
Cu	100.0 ± 0.6	3114	121207	Se	100 ± 1	3149	100901
Dy	100.0 ± 0.6	3115a	990504	Na	100.0 ± 0.6	3152a	120715
Er	100.0 ± 0.6	3116a	000831	Sr	100.0 ± 0.6	3153a	990906
Eu	100.0 ± 0.6	3117a	120705	Tb	100.0 ± 0.6	3157a	100518
Gd	100.0 ± 0.6	3118a	992004	Tl	100.0 ± 0.6	3158	993012
Ga	100.0 ± 0.6	3119a	890709	Th	100 ± 1	*	
Ho	100.0 ± 0.6	3123a	090408	Tm	100 ± 1	3160a	790912
In	100.0 ± 0.6	3124a	110516	U	100.0 ± 0.6	3164	080521
Fe	100.0 ± 0.6	3126a	140812	V	100 ± 1	3165	992706
La	100.0 ± 0.6	3127a	890402	Yb	100.0 ± 0.6	3166a	790512
Pb	100 ± 1	3128	101026	Y	100.0 ± 0.6	3167a	120314
Li	100 ± 1	3129a	100714	Zn	100.0 ± 0.6	3168a	120629

The Certified values are based on gravimetric and volumetric preparation, and verified against SRM 3100 series via inductively coupled plasma optical emission spectrometry (ICP-OES) using an internal laboratory developed method. The uncertainty in the certified value is calculated for a 95% confidence interval and coverage factor *k* is about 2.

- Refer to Traceability Information, Section d

## Preparation Information:

The standard is generally manufactured from CRM single element solutions that were prepared using the methods developed at NIST for SRM Spectrometric Standard Solutions under appropriate laboratory conditions. Sub-boiling distilled high-purity acid has been used to place the materials in solution and to stabilize the standard. The matrix is as noted above in 18 megaohm deionized water.

Stability of this product is based upon rigorous short term and long term testing of the solution for the certified value. This testing includes, but is not limited to, the effect of temperature and packaging on the product.

## Intended Use:

Lot No.: SAMPLE

Rev. No.: 4.1.0

Page 1 of 2

High-Purity Standards is certified to ISO 9001:2008 and accredited to ISO/IEC 17025:2005 and ISO Guide 34:2009.

The CRM is intended for use as a calibration standard and QC control for instruments such as ICPOES, ICPMS, AAS and XRF, and validation of analytical methods. It also can be used in EPA, ASTM and other methods.

#### Traceability Information:

The traceability of this standard is maintained through an unbroken chain of comparisons to appropriate standards with suitable procedure and measurement uncertainties. The maintenance of the base and derived units of International System of Units (SI) with traceability of measurement results (contemporary metrology) to SI ensures their comparability over time as follows.

a. **Standard Weight and Analytical Balance**

The standard weights (NBS weights Inventory No 20231A) are calibrated every two years by South Carolina Metrology Laboratory that is a participant in "NIST Weights and Measures Measurement Assurance Program" with a certificate of measurement traceability to NIST primary standards.

The balances are calibrated yearly by the ISO 17025 accredited metrology service, and are verified weekly by an in-house method using standard weights.

b. **Volumetric Device**

The calibration of volumetric vessels is checked annually using the ASTM method E542.

c. **Thermometer**

The standard thermometers are calibrated every year by the ISO 17025 accredited metrology service. The thermometers used in-house are verified against the standard thermometers yearly.

d. **Calibration Standards**

The Calibration Standards are traceable to SRM 3100 Series Spectrometric Standard Solutions or second sources.

#### Packaging and Storage Conditions:

The standard is packaged in a pre-cleaned polyethylene bottle. To maintain the integrity of this product, the solution should be kept tightly capped and stored under normal laboratory conditions.

#### Refer to Material Safety Datasheet (MSDS) for hazardous information.

#### Expiration Information:

The expiry date is guaranteed to be valid for twelve months from the shipping date provided. For this reason, standards from the same lot may have different expiration dates.

**Preparation Date:** June 5, 2015

**Shipped Date:**

**Expiration Date:**

**Certificate Issue Date:** June 15, 2015

#### Quality Information:



ISO/IEC 17025:2005 Accreditation  
Certificate Number AT-1529



ISO Guide 34:2009 (RMP) Accreditation  
Certificate Number AR-1436



**Angel Sellers**  
**Quality Manager**

NOTICE: HPS products are intended for laboratory use only. All products should be handled and used by trained professional personnel. The responsibility for the safe handling and use of these products rests solely with the buyer and/or user. The data and information as stated was furnished by the manufacturer of the product. The information provided in this certificate pertains only to the lot number specified. None of the information provided in this certificate may be used, reproduced or transmitted in any form or by any means without written approval from High Purity Standards.

Lot No.: SAMPLE

Rev. No.: 4.1.0

Page 2 of 2

**High-Purity Standards is certified to ISO 9001:2008 and accredited to ISO/IEC 17025:2005 and ISO Guide 34:2009.**