

# Certificate of Analysis

## Product Description:

Part Number: **ICP-MS-SS-2**  
Lot Number: **SAMPLE**  
Matrix: **2% HNO<sub>3</sub>**  
Purity: **99.964% - 99.9999%**

## Certified Values:

Element	( $\mu\text{g/L}$ )	SRM ID	SRM Lot#
Ag	6.0 $\pm$ 0.3	3151	992212
Al	10.0 $\pm$ 0.5	3101a	060502
Ba	4.0 $\pm$ 0.2	*	
Be	35.00 $\pm$ 1.75	3105a	090514
Bi	3.00 $\pm$ 0.15	3106	991212
Ce	3.00 $\pm$ 0.15	3110	090504
Co	8.0 $\pm$ 0.4	3113	000630
Cs	3.00 $\pm$ 0.15	3111a	130228
Cu	15.00 $\pm$ 0.75	3114	121207
Ga	10.0 $\pm$ 0.5	3119a	890709
Ho	3.00 $\pm$ 0.15	3123a	090408
In	3.00 $\pm$ 0.15	3124a	110516
Li	8.0 $\pm$ 0.4	3129a	100714
Mg	10.0 $\pm$ 0.5	3131a	140110
Mn	6.0 $\pm$ 0.3	3132	050429
Ni	15.00 $\pm$ 0.75	3136	120619
Rh	3.00 $\pm$ 0.15	3144	070619
Sc	8.0 $\pm$ 0.4	3148a	100701
Sr	5.00 $\pm$ 0.25	3153a	990906
Ta	3.00 $\pm$ 0.15	3155	110715
Tb	3.00 $\pm$ 0.15	3157a	100518
Tl	4.0 $\pm$ 0.2	3158	993012
V	3.00 $\pm$ 0.15	3165	992706
Y	3.00 $\pm$ 0.15	3167a	120314
Zn	20 $\pm$ 1	3168a	120629

The Certified values are based on gravimetric and volumetric preparation, and verified against SRM 3100 series developed by National Institute of Standards and Technology (NIST) via inductively coupled plasma optical

Lot No.: SAMPLE

Rev. No.: 3.2.0

Page 1 of 3

High-Purity Standards is certified to ISO 9001:2008 and accredited to ISO/IEC 17025:2005 and ISO Guide 34:2009.

emission spectrometry (ICP-OES) and inductively coupled plasma mass spectrometry (ICP-MS) using an internal laboratory developed method. The uncertainty in the certified value is calculated for a 95% confidence interval and coverage factor  $k$  is about 2

\* Refer to Traceability Information, Section d

### Preparation Information:

The standard is generally prepared from single element standard solutions that are ISO Guide 34 certified reference materials. Highest purity source materials were purchased from qualified vendors per ISO 9001:2008 guidelines and assayed by ICP-OES for conformity prior to use. Sub-boiling distilled high-purity acid has been used to place the materials in solution and to stabilize the standard. The matrix is as noted above in 18 megaohm deionized water.

### Traceability Information:

The traceability of this standard is maintained through an unbroken chain of comparisons to appropriate standards with suitable procedure and measurement uncertainties. The maintenance of the base and derived units of International System of Units (SI) with traceability of measurement results (contemporary metrology) to SI ensures their comparability over time as follows.

a. **Standard Weight and Analytical Balance**

The standard weights (NBS weights Inventory No 20231A) are calibrated every two years by South Carolina Metrology Laboratory that is a participant in "NIST Weights and Measures Measurement Assurance Program" with a certificate of measurement traceability to NIST primary standards.

The balances are calibrated yearly by the ISO 17025 accredited metrology service, and are verified weekly by an in-house method using standard weights.

b. **Volumetric Device**

The calibration of volumetric vessels is checked annually using the ASTM method E542

c. **Thermometer**

The standard thermometers are calibrated every year by the ISO 17025 accredited metrology service. The thermometers used in-house are verified against the standard thermometers yearly.

d. **Calibration Standards**

The Calibration Standards are traceable to SRM 3100 Series Spectrometric Standard Solutions or second sources.

### Packaging and Storage Conditions:

The standard is packaged in a pre-cleaned polyethylene bottle. To maintain the integrity of this product, the solution should be kept tightly capped and stored under normal laboratory conditions.

### Refer to Material Safety Datasheet (MSDS) for hazardous information.

### Expiration Information:

The expiry date is guaranteed to be valid for six months from the shipping date provided. For this reason, standards from the same lot may have different expiration dates.

**Preparation Date:** June 9, 2015

**Shipped Date:**

**Expiration Date:**

**Certificate Issue Date:** June 9, 2015

## Quality Information:



**ISO/IEC 17025:2005 Accreditation**  
**Certificate Number AT-1529**



**Angel Sellers**  
**Quality Manager**

NOTICE: HPS products are intended for laboratory use only. All products should be handled and used by trained professional personnel. The responsibility for the safe handling and use of these products rests solely with the buyer and/or user. The data and information as stated was furnished by the manufacturer of the product. The information provided in this certificate pertains only to the lot number specified. None of the information provided in this certificate may be used, reproduced or transmitted in any form or by any means without written approval from High Purity Standards.

SAMPLE