

Certificate of Analysis

Product Description:

Name: ICP-MS Calibration Standard
Part Number: **ICP-MSCS**
Lot Number: **SAMPLE**
Matrix: 2% HNO₃ + Tr HF
Purity: 99.99% - 99.9999%

Certified Values:

Element	($\mu\text{g/mL}$)	SRM ID	SRM Lot#	Element	($\mu\text{g/mL}$)	SRM ID	SRM Lot#
Al	10.00 \pm 0.06	3101a	060502	Mg	10.00 \pm 0.06	3131a	050302
Sb	10.0 \pm 0.1	*		Mn	10.00 \pm 0.15	3132	050429
As	10.0 \pm 0.1	3103a	100818	Mo	10.0 \pm 0.1	3134	891307
Ba	10.00 \pm 0.06	3104a	070222	Ni	10.00 \pm 0.06	3136	120619
Be	10.0 \pm 0.1	3105a	090514	Sc	10.0 \pm 0.1	3148a	100701
B	10.0 \pm 0.1	3107	070514	Se	10.0 \pm 0.1	3149	100901
Cd	10.00 \pm 0.06	3108	130116	Ag	10.00 \pm 0.06	3151	992212
Ca	10.00 \pm 0.06	3109a	050825	Na	10.00 \pm 0.06	3152a	120715
Cr	10.00 \pm 0.06	3112a	030730	Sr	10.00 \pm 0.06	3153a	990906
Co	10.00 \pm 0.06	3113	000630	Tl	10.00 \pm 0.06	3158	993012
Cu	10.00 \pm 0.06	3114	121207	Th	10.0 \pm 0.1	*	
Eu	10.00 \pm 0.06	3117a	120705	U	10.00 \pm 0.06	3164	080521
Ho	10.00 \pm 0.06	3123a	090408	V	10.0 \pm 0.1	3165	992706
La	10.00 \pm 0.06	3127a	890402	Yb	10.00 \pm 0.06	3166a	790512
Pb	10.0 \pm 0.1	3128	101026	Zn	10.00 \pm 0.06	3168a	120629
Li	10.0 \pm 0.1	3129a	100714				

The Certified values are based on gravimetric and volumetric preparation, and verified against SRM 3100 series via inductively coupled plasma optical emission spectrometry (ICP-OES) using an internal laboratory developed method. The uncertainty in the certified value is calculated for a 95% confidence interval and coverage factor k is about 2.

* Refer to Traceability Information, Section d

Preparation Information:

The standard is generally manufactured from CRM single element solutions that were prepared using the methods developed at NIST for SRM Spectrometric Standard Solutions under appropriate laboratory conditions. Sub-boiling distilled high-purity acid has been used to place the materials in solution and to stabilize the standard. The matrix is as noted above in 18 megaohm deionized water.

Stability of this product is based upon rigorous short term and long term testing of the solution for the certified value. This testing includes, but is not limited to, the effect of temperature and packaging on the product.

Intended Use:

The CRM is intended for use as a calibration standard and QC control for instruments such as ICPOES, ICPMS, AAS and XRF, and validation of analytical methods. It also can be used in EPA, ASTM and other methods.

Traceability Information:

Lot No.: SAMPLE

Rev. No.: 4.1.0

Page 1 of 2

High-Purity Standards is certified to ISO 9001:2008 and accredited to ISO/IEC 17025:2005 and ISO Guide 34:2009.

The traceability of this standard is maintained through an unbroken chain of comparisons to appropriate standards with suitable procedure and measurement uncertainties. The maintenance of the base and derived units of International System of Units (SI) with traceability of measurement results (contemporary metrology) to SI ensures their comparability over time as follows.

a. **Standard Weight and Analytical Balance**

The standard weights (NBS weights Inventory No 20231A) are calibrated every two years by South Carolina Metrology Laboratory that is a participant in "NIST Weights and Measures Measurement Assurance Program" with a certificate of measurement traceability to NIST primary standards.

The balances are calibrated yearly by the ISO 17025 accredited metrology service, and are verified weekly by an in-house method using standard weights.

b. **Volumetric Device**

The calibration of volumetric vessels is checked annually using the ASTM method E542.

c. **Thermometer**

The standard thermometers are calibrated every year by the ISO 17025 accredited metrology service. The thermometers used in-house are verified against the standard thermometers yearly.

d. **Calibration Standards**

The Calibration Standards are traceable to SRM 3100 Series Spectrometric Standard Solutions or second sources.

Packaging and Storage Conditions:

The standard is packaged in a pre-cleaned polyethylene bottle. To maintain the integrity of this product, the solution should be kept tightly capped and stored under normal laboratory conditions.

Refer to Material Safety Datasheet (MSDS) for hazardous information.

Expiration Information:

The expiry date is guaranteed to be valid for twelve months from the shipping date provided. For this reason, standards from the same lot may have different expiration dates.

Preparation Date: December 8, 2014

Shipped Date:

Expiration Date:

Certificate Issue Date: December 23, 2014

Quality Information:



**ISO/IEC 17025:2005 Accreditation
Certificate Number AT-1529**



**ISO Guide 34:2009 (RMP) Accreditation
Certificate Number AR-1436**



**Angel Sellers
Quality Manager**

NOTICE: HPS products are intended for laboratory use only. All products should be handled and used by trained professional personnel. The responsibility for the safe handling and use of these products rests solely with the buyer and/or user. The data and information as stated was furnished by the manufacturer of the product. The information provided in this certificate pertains only to the lot number specified. None of the information provided in this certificate may be used, reproduced or transmitted in any form or by any means without written approval from High Purity Standards.